



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L3631-260213)

This is to certify that

MPI Corporation **MPI Testing Laboratory**

Rm. 7, 2 F., No. 1, Huanke 1st Rd., Zhubei City, Hsinchu County 302047, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 3631

Originally Accredited : December 20, 2019

Effective Period : December 20, 2025 to December 19, 2028

Accredited Scope : Testing Field, see described in the Appendix



Scan to verify

Yi-Ling Chen

Yi-Ling Chen
President, Taiwan Accreditation Foundation
February 13, 2026

Accreditation Number : 3631

Laboratory Head : WEI, Hao

▼ 19. 02 Electronic and Electric

Electric & electronic products (include components)

E025 Insertion loss

EIA-364-101

Frequency Range: 50 MHz~500 MHz, Measurement uncertainty: 0.78 dB

Frequency Range: 500 MHz~2 GHz, Measurement uncertainty: 0.24 dB

Frequency Range: 2 GHz~26.5 GHz, Measurement uncertainty: 0.30 dB

Frequency Range: 26.5 GHz~40 GHz, Measurement uncertainty: 0.42 dB

Approval Signatory: CHOU, Chia-Nan; WEI, Hao

E025 Characteristic Impedance

EIA-364-108

EIA-364-108A

Bandwidth: 20 GHz (max.)

Reflected rise time: 30 ps (min.)

Approval Signatory: CHOU, Chia-Nan; WEI, Hao

(Null below)

